

Section 5 – 510(k) Summary

JUL 13 2012

Date Prepared: March 16, 2012

Company: Angiotech
100 Dennis Dr.
Reading, PA 19606

Contact: Kirsten Stowell
Regulatory Affairs Manager
Phone: 610-404-3367
Fax: 610-404-3924
Email: kstowell@angio.com

Device trade name: Quill™ PDO Knotless Tissue-Closure Device (Polydioxanone)

Device Common Name: Polydioxanone Absorbable Surgical Suture

Device classification: Absorbable Polydioxanone Surgical Suture
Product code, NEW
21 CFR 878.4840
Class II

Legally marketed devices to which the device is substantially equivalent:

K051609	Quill Synthetic Absorbable Barbed Suture
K080680	
K080985	
N18331	Ethicon PDS II™ (Polydioxanone) Sutures

Description of the device: The Quill™ Knotless Tissue-Closure Device comprised of PDO (Polydioxanone) is a synthetic absorbable tissue-closure device that is intended for use in the closure of soft tissue. It is comprised of polyester [poly (p-dioxanone)] monofilament suture material, dyed with D&C Violet No. 2. The device is designed with small bi-directional barbs along the length that eliminate the need to tie knots during approximation. The device is available in various lengths in diameter sizes 2 through 4-0 with various needles attached to each end.

Indications for Use: Quill™ PDO Knotless Tissue-Closure Device, comprised of Polydioxanone, is indicated for soft tissue approximation where use of an absorbable suture is appropriate.

**Substantial
Equivalence:**

The Quill™ Knotless Tissue-Closure Device comprised of PDO (Polydioxanone) has the same design and materials as its predicates, including the same intended use, technological characteristics and size ranges as the predicate devices.

The intended use is identical to both predicate devices. The design, material and chemical composition of the proposed device is identical to the predicate Quill Synthetic Absorbable Barbed Suture (K051609, K080680, K080985).

Performance tests:

Pre-clinical animal performance testing was conducted to confirm and support the information provided in the Quill™ Knotless Tissue-Closure Device comprised of PDO (Polydioxanone) Instructions for Use.

The results of this testing demonstrates that the Quill™ Knotless Tissue-Closure Device comprised of PDO (Polydioxanone) is substantially equivalent in safety and performance to the predicate devices.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room -WO66-G609
Silver Spring, MD 20993-0002

Angiotech
% Ms. Kirsten Stowell
Regulatory Affairs Manager
100 Dennis Drive
Reading, Pennsylvania 19606

JUL 13 2012

Re: K120827

Trade/Device Name: Quill™ Knotless Tissue-Closure Device, Polydioxanone
Regulation Number: 21 CFR 878.4840
Regulation Name: Absorbable polydioxanone surgical suture
Regulatory Class: II
Product Code: NEW
Dated: July 02, 2012
Received: July 05, 2012

Dear Ms. Stowell:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21


Page 2 - Ms. Kirsten Stowell

CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucml15809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Mark N. Melkerson
Director
Division of Surgical, Orthopedic
and Restorative Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Section 4 – Indications for Use Statement

510k number if known: K120827

Device Name: Quill™ PDO Knotless Tissue-Closure Device, Polydioxanone

Indications for Use:

Quill™ PDO Knotless Tissue-Closure Device, comprised of Polydioxanone, is indicated for soft tissue approximation where use of an absorbable suture is appropriate.

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE
IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

David Kline for MM
(Division Sign-Off)
Division of Surgical, Orthopedic,
and Restorative Devices

510(k) Number K120827